

***AAMI Consensus Report***

End User Disclosures  
for Emergency Use

**Resuscitator Systems**

*AAMI/CR504:2020*



# End user disclosures for emergency use resuscitator systems (EURS)

Revisions are expected to be made to this document as the COVID-19 situation evolves. Please go to [https://www.aami.org/covid\\_cr](https://www.aami.org/covid_cr) to find the most current version as well as past versions. This document is freely available and may be shared with all interested stakeholders. Contact [celliot@ami.org](mailto:celliot@ami.org) with any comments or questions.

Approved 8 April 2020 by  
AAMI

**Abstract:** Identifies high priority hazards and their causes to be considered in development and the information to be disclosed by emergency use resuscitator system (EURS) manufacturers to the end user. These are based on the hazards identified in IEC 60601-1 and ISO 10651-4.

**Keywords:** COVID-19

## AAMI Consensus Report

A Consensus Report (CR) is a publication of the Association for the Advancement of Medical Instrumentation (AAMI) developed to provide concise, prompt and practical guidance on narrowly focused topics of high importance to the health technology community. A Consensus Report is intended provide initial consensus guidance in response to an urgent/immediate need for guidance in the following instances:

- While more robust data/information develops on emergent areas
- When variation in the development, implementation or use of a product or process exists
- When existing standards or other documents require additional context/clarification

A Consensus Report is not subject to the same formal process as a standard and while similar in nature to a technical information report (TIR), a CR is based on the collective knowledge and experience of a selected group of stakeholders and has not undergone the wider reviews of a TIR or standard and offers an even greater response time.

**CAUTION NOTICE:** This AAMI CR may be revised or withdrawn at any time. Because it addresses a rapidly evolving field or technology, readers are cautioned to ensure that they have also considered information that may be more recent than this document.

All standards, technical information reports, consensus reports and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Comments on this technical information report are invited and should be sent to AAMI, Attn: Standards Department, 901 N. Glebe Rd, Suite 300, Arlington, VA 22203-1633.

### *Published by*

AAMI  
901 N. Glebe Rd., Suite 300  
Arlington, VA 22203-1633  
[www.aami.org](http://www.aami.org)

© 2020 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, visit the [Copyright Clearing Center](#)

Printed in the United States of America

**ISBN 978-1-57020-751-8**

## Contents

Page

Task Group representation .....	iv
Acknowledgments .....	v
Purpose.....	1
1 Electrical Shock Hazard.....	1
2 Mechanical Hazards .....	2
3 Environmental Hazards .....	2
4 CO <sub>2</sub> Rebreathing .....	2
5 Reuse Hazards.....	3
6 Biocompatibility .....	3
7 Electromagnetic Compatibility (EMC).....	3
8 Alarm System .....	3
9 Accuracy of controls .....	4
10 Accessories.....	4
11 Programmable Electrical Medical Systems.....	4
12 Risk Management Process .....	4
13 Other hazards.....	5

## Task Group representation

### Association for the Advancement of Medical Instrumentation

#### COVID-19 Response Team Members

This AAMI Consensus Report (CR) was developed by a task group under the auspices of the AAMI COVID-19 Response Team.

The **AAMI COVID-19 Response Team** had the following members:

*Cochairs:* Jennifer Danieley  
David Feinstein  
Julian Goldman

*Members:* Simona Bancos, FDA/CDRH  
Andrew Bath, ResMed Inc.  
Brandon Blakely, FDA/CDRH  
Brad Bonnette, ECRI Institute  
Caitlin Brady, Intertek  
David Busch, UT Southwestern Medical Center  
Anthony Ciccarello, Philips  
Steven Dain, University of Western Ontario  
Rakhi Dalal, FDA/CDRH  
Jennifer Danieley, FDA/CDRH  
Andy Doering, Medtronic  
Simon Dunham, Weill Cornell Medicine  
David Feinstein, American Society of Anesthesiologists (ASA)  
Bruce Friedman, GE Healthcare  
Hamed Ghods, FDA/CDRH  
Julian Goldman, Partners HealthCare System  
Ralf Heesch, Draeger Medical Systems Inc.  
Heidi Horn, Nuvolo Technologies  
Fernando Isaza, Philips  
Michael Jaffe, Cardiorespiratory Consulting LLC  
Gardner Kimm, Medtronic Inc Campus  
Robert Kopotic, Edwards Lifesciences  
Hubertus Lasthaus, VitalAire Germany  
Ed Madsen, Avanos Medical  
Phoebe Mainland, Alfred Health  
Madeleine Manousaridis, Standards Australia  
Benoit Marchal, Air Liquide  
Thomas Marmet, GE Healthcare  
Debra Milamed, Harvard University  
Cyndy Miller, Medtronic Inc Campus  
Bryant Moeller, ResMed Inc.  
Curtis Morgan, 3M Health Care  
Akito Ohmura, Teikyo University-Mizonokuchi Hospital  
David Osborn, Philips  
John Stark, 3M Health Care  
Robert Steurer, Steurer Consulting Group  
Dongbo Wang, FDA/CDRH  
Sandy Weininger, FDA/CDRH

---

NOTE—Participation by federal agency representatives in the development of this document does not constitute endorsement by the federal government or any of its agencies.

---

## **Acknowledgments**

AAMI gratefully acknowledges the writing team members, Julian Goldman, Dave Osborn, Anthony Ciccarello and Sandy Weinger for their outstanding and expeditious work in preparing these drafts for committee review and approval.





# End user disclosures for emergency use resuscitator systems (EURS)

## Purpose

The goal of this document is to identify high priority hazards and their causes to be considered in development and the information to be disclosed by emergency use resuscitator system (EURS) manufacturers to the end user. These are based on the hazards identified in IEC 60601-1<sup>1</sup> and ISO 10651-4<sup>2</sup>.

NOTE This document is intended to be used in conjunction with AAMI CR503:2020, *Emergency use resuscitator systems (EURS) design guidance*.

## 1 Electrical Shock Hazard

Purpose: to ensure adequate patient and operator safety in terms of shock (leakage current, dielectric strength, ground continuity).

### Disclosures:

- List AC input power requirements of the EURS (voltage, frequency, amperes).
- DC power input requirement, if applicable.
- Indicate the electrical classification of EURS:
  - Class I (EURS has a protective earth connection with a 3-wire power cord)
  - Class II (EURS does not have a protective earth ground but is double insulated with a 2-wire power cord)
  - Internally powered (powered by a rechargeable battery inside the EURS or external to EURS)

NOTE An EURS can have more than one classification e.g., Class II/internally powered.

- If the power supply connected to mains power is not medical grade (i.e., IEC 60601-1 compliant), describe the means used to reduce leakage currents to IEC 60601-1 limits (e.g. use of an isolation transformer, second permanently installed protective earth connection).
- If the power supply connected to mains power is Class I, add a warning:

<sup>1</sup> IEC 60601-1, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

<sup>2</sup> ISO 10651-4, *Lung ventilators – Part 4: Particular requirements for operator powered resuscitators*

27           Warning: This ventilator relies on the integrity of the protective earth ground to reduce the risk of  
28           electrical shock. Check the integrity and verify the function of the protective earth ground of the  
29           supply mains receptacle prior to use.

- 30   • Describe the type of patient connection: basic, basic floating, cardiac floating (type B, BF or CF) and  
31   defibrillation-proof.

## 32   **2       Mechanical Hazards**

- 33   a) Purpose: to ensure that the EURS can withstand mechanical stresses from being carried or wheeled  
34   while being transported indoors or outdoors.

### 35   Disclosures:

- 36   • Identify the mobility of the EURS:
  - 37       ○ Transit operable: EURS is intended to operate while being moved.
  - 38       ○ Portable: EURS is intended to be carried (but not operating) from one location to another.
  - 39       ○ Mobile: EURS is intended to be wheeled (but not operating) from one location to another.
- 40   b) Purpose: to ensure that the moving parts of the EURS do not pose an unacceptable risk to the patient  
41   or operator.

### 42   Disclosures:

- 43   • If the EURS has wheels, assess the stability and disclose the safe angle before tipping occurs.
- 44   • Identify any trapping zones (e.g. trapping fingers, hair, PPE) and how they are guarded.

## 45   **3       Environmental Hazards**

46   Purpose: to ensure that the EURS can be stored and operated in its intended environment.

### 47   Disclosures:

- 48   • Indicate the temperature/humidity/altitude range over which the EURS is intended to operate and  
49   meets its specifications.
- 50   • Indicate the intended range of conditions (temperature/humidity specifications) in which the EURS  
51   can be stored.

## 52   **4       CO<sub>2</sub> Rebreathing**

53   Purpose: to reduce the risk of excessive carbon dioxide in the bloodstream.

### 54   Disclosures:

- 55   • Describe the means implemented to minimize the risk of rebreathing and to keep residual exhaled  
56   CO<sub>2</sub> to acceptable levels.

57 **5 Reuse Hazards**

58 Purpose: to reduce the risk of cross contamination.

59 Disclosures:

- 60 • Describe the cleaning and disinfection procedures needed between uses and between patients for  
61 both the EURS and the accessories.
- 62 • Description of location and specifications of required EURS particle filters and replacement  
63 intervals.

64 **6 Biocompatibility**

65 Purpose: to reduce the risk of biological reaction to foreign substances.

66 Disclosures:

- 67 • For the gas pathway, indicate if any biocompatibility evaluations were performed per ISO 18562  
68 (series)<sup>3</sup>.
- 69 • For parts intended to touch the patient, indicate if any biocompatibility evaluations were performed  
70 per ISO 10993 (series)<sup>4</sup>.

71 **7 Electromagnetic Compatibility (EMC)**

72 Purpose: to ensure that the EURS is adequately protected from electromagnetic emissions from other  
73 electrical sources (e.g. cell phones, ESD) and to ensure that the EURS does not interfere with the operation  
74 of other nearby electronic medical devices.

75 Disclosures:

- 76 • Indicate if any EMC testing was performed and identify the standards (e.g., IEC 60601-1-2<sup>5</sup>) to  
77 which the EURS was evaluated.
- 78 • If EMC testing has not been performed, add a warning:
- 79 This ventilator has not been tested for electromagnetic compatibility (EMC). It may produce  
80 electromagnetic disturbances that will affect the performance of other equipment. It may fail to  
81 perform as expected in the presence of electromagnetic disturbances from other equipment.

82 **8 Alarm System**

83 Purpose: to reduce the risk to the patient by alerting the caregiver of a hazardous situation.

---

<sup>3</sup> ISO 18562, *Biocompatibility evaluation of breathing gas pathways in healthcare applications*

<sup>4</sup> ISO 10993, *Biological evaluation of medical devices*

<sup>5</sup> IEC 60601-1-2, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests*

84 Disclosures:

- 85 • Describe the functionality of the alarm system.
- 86 • List available alarm conditions, their relative priority and default alarm limits.
- 87 • Describe the default alarm settings (e.g. latched, not latched alarm signals, alarm condition  
88 disabled).
- 89 • Indicate the means by which the auditory alarm signal can be inactivated and for how long.

## 90 **9 Accuracy of controls**

91 Purpose: to reduce the risk of hazardous output from the EURS to the patient.

92 Disclosures:

- 93 • List of displayed parameters: e.g., pressure, tidal volume, respiratory rate.
- 94 • Describe how the displayed parameters are measured or determined.
- 95 • List the accuracy of therapy parameters.

## 96 **10 Accessories**

97 Purpose: to ensure the safe use of the EURS with compatible accessories

98 Disclosures:

- 99 • List of recommended accessories and their replacement intervals e.g. tubing, patient interface,  
100 filters, replacement batteries.

## 101 **11 Programmable Electrical Medical Systems**

102 Purpose: to ensure that the software operates safely and as specified.

103 Disclosures:

- 104 • Indicate whether the software was developed under a controlled life cycle process (e.g.,  
105 IEC 62304<sup>6</sup>).
- 106 • List any known unresolved software anomalies and workarounds.
- 107 • Indicate: Due to the rapid development cycle for this emergency use device, all efforts were made  
108 to verify the software, but defects may still exist. The consequences of these defects are unknown  
109 and may pose a risk to the patient.

## 110 **12 Risk Management Process**

111 Purpose: to ensure risks were comprehensively identified and adequately managed.

---

<sup>6</sup> IEC 62304, *Medical device software — Software life cycle processes*

112 Disclosures:

- 113 • Indicate whether the EURS design has been developed using a risk management process (e.g.,  
114 ISO 14971<sup>7</sup>).

### 115 **13 Other hazards**

116 Purpose: to reduce the risk of thermal injury or other events.

117 Disclosures:

- 118 • If applicable, indicate the battery specifications including:
- 119 ○ the type of battery and chemistry;
- 120 ○ a description of the means to determine the status of the battery (e.g., charging, low battery  
121 indicator);
- 122 ○ conformance to applicable standards (e.g., IEC 62133<sup>8</sup> for rechargeable batteries or IEC  
123 60086-4<sup>9</sup> for non-rechargeable batteries).
- 124 • Indicate the ingress protection (IP) of the EURS enclosure: IP 22 is recommended (protection  
125 against foreign objects  $\geq 12.5$  mm and against dripping (15° tilted) water).
- 126 • Indicate if the EURS is suitable for use in an oxygen enriched environment  $> 25\%$  O<sub>2</sub> (are adequate  
127 protections in place to reduce risk of fire ignition).
- 128 • If the EURS contains oxygen at pressures exceeding 5 bar, the protections taken to ensure that  
129 auto-ignition from adiabatic compression cannot occur (e.g., parts of the EURS operating at  
130 pipeline pressure).

---

<sup>7</sup> ISO 14971, *Medical devices - Application of risk management to medical devices*

<sup>8</sup> IEC 62133, *Secondary cells and batteries containing alkaline or other non-acid electrolytes – Safety requirements for portable sealed secondary cells, and for batteries made from them, for use in portable applications*

<sup>9</sup> IEC 60086-4, *Primary batteries – Part 4: Safety of lithium batteries*