

AAMI Consensus Report

End User Disclosures for
Emergency Use Ventilators
(EUVs)

AAMI/CR502:2020

End user disclosures for emergency use ventilators (EUVs)

Revisions are expected to be made to this document as the COVID-19 situation evolves. Please go to https://www.aami.org/covid_cr to find the most current version as well as past versions. This document is freely available and may be shared with all interested stakeholders. Contact celliot@ami.org with any comments or questions.

Approved 8 April 2020 by
AAMI

Abstract: Identifies high priority hazards and their causes to be considered in development and the information to be disclosed by Emergency Use Ventilator (EUV) manufacturers to the end user. These are based on the hazards identified in IEC 60601-1 and ISO 80601-2-80.

Keywords: COVID-19

AAMI Consensus Report

A Consensus Report (CR) is a publication of the Association for the Advancement of Medical Instrumentation (AAMI) developed to provide concise, prompt and practical guidance on narrowly focused topics of high importance to the health technology community. A Consensus Report is intended provide initial consensus guidance in response to an urgent/immediate need for guidance in the following instances:

- While more robust data/information develops on emergent areas
- When variation in the development, implementation or use of a product or process exists
- When existing standards or other documents require additional context/clarification

A Consensus Report is not subject to the same formal process as a standard and while similar in nature to a technical information report (TIR), a CR is based on the collective knowledge and experience of a selected group of stakeholders and has not undergone the wider reviews of a TIR or standard and offers an even greater response time.

CAUTION NOTICE: This AAMI CR may be revised or withdrawn at any time. Because it addresses a rapidly evolving field or technology, readers are cautioned to ensure that they have also considered information that may be more recent than this document.

All standards, technical information reports, consensus reports and other types of technical documents developed by AAMI are *voluntary*, and their application is solely within the discretion and professional judgment of the user of the document. Occasionally, voluntary technical documents are adopted by government regulatory agencies or procurement authorities, in which case the adopting agency is responsible for enforcement of its rules and regulations.

Comments on this technical information report are invited and should be sent to AAMI, Attn: Standards Department, 901 N. Glebe Rd, Suite 300, Arlington, VA 22203-1633.

Published by

AAMI
901 N. Glebe Rd., Suite 300
Arlington, VA 22203-1633
www.aami.org

© 2020 by the Association for the Advancement of Medical Instrumentation

All Rights Reserved

Publication, reproduction, photocopying, storage, or transmission, electronically or otherwise, of all or any part of this document without the prior written permission of the Association for the Advancement of Medical Instrumentation is strictly prohibited by law. It is illegal under federal law (17 U.S.C. § 101, *et seq.*) to make copies of all or any part of this document (whether internally or externally) without the prior written permission of the Association for the Advancement of Medical Instrumentation. Violators risk legal action, including civil and criminal penalties, and damages of \$100,000 per offense. For permission regarding the use of all or any part of this document, visit the [Copyright Clearing Center](http://www.copyright.com)

Printed in the United States of America

ISBN 978-1-57020-749-5

Contents

Page

Task Group representation	iv
Acknowledgments	v
Purpose.....	1
1 Electrical Shock Hazard.....	1
2 Mechanical Hazards	2
3 Environmental Hazards	2
4 Reuse Hazards.....	2
5 Biocompatibility	3
6 Electromagnetic Compatibility (EMC).....	3
7 Alarm System.....	3
8 Accuracy of controls	4
9 Accessories.....	4
10 Programmable Electrical Medical Systems	4
11 Risk Management Process.....	4
12 Other hazards	4

Task Group representation

Association for the Advancement of Medical Instrumentation

COVID-19 Response Team Members

This AAMI Consensus Report (CR) was developed by a task group under the auspices of the AAMI COVID-19 Response Team.

The **AAMI COVID-19 Response Team** had the following members:

Cochairs: Jennifer Danieley
David Feinstein
Julian Goldman

Members: Simona Bancos, FDA/CDRH
Andrew Bath, ResMed Inc.
Brandon Blakely, FDA/CDRH
Brad Bonnette, ECRI Institute
Caitlin Brady, Intertek
David Busch, UT Southwestern Medical Center
Anthony Ciccarello, Philips
Steven Dain, University of Western Ontario
Rakhi Dalal, FDA/CDRH
Jennifer Danieley, FDA/CDRH
Andy Doering, Medtronic
Simon Dunham, Weill Cornell Medicine
David Feinstein, American Society of Anesthesiologists (ASA)
Bruce Friedman, GE Healthcare
Hamed Ghods, FDA/CDRH
Julian Goldman, Partners HealthCare System
Ralf Heesch, Draeger Medical Systems Inc.
Heidi Horn, Nuvolo Technologies
Fernando Isaza, Philips
Michael Jaffe, Cardiorespiratory Consulting LLC
Gardner Kimm, Medtronic Inc Campus
Robert Kopotic, Edwards Lifesciences
Hubertus Lasthaus, VitalAire Germany
Ed Madsen, Avanos Medical
Phoebe Mainland, Alfred Health
Madeleine Manousaridis, Standards Australia
Benoit Marchal, Air Liquide
Thomas Marmet, GE Healthcare
Debra Milamed, Harvard University
Cyndy Miller, Medtronic Inc Campus
Bryant Moeller, ResMed Inc.
Curtis Morgan, 3M Health Care
Akito Ohmura, Teikyo University-Mizonokuchi Hospital
David Osborn, Philips
John Stark, 3M Health Care
Robert Steurer, Steurer Consulting Group
Dongbo Wang, FDA/CDRH
Sandy Weininger, FDA/CDRH

NOTE—Participation by federal agency representatives in the development of this document does not constitute endorsement by the federal government or any of its agencies.

Acknowledgments

AAMI gratefully acknowledges the writing team members, Julian Goldman, Dave Osborn, Anthony Ciccarello and Sandy Weinger for their outstanding and expeditious work in preparing these drafts for committee review and approval.

End user disclosures for emergency use ventilators (EUVs)

Purpose

The goal of this document is to identify high priority hazards and their causes to be considered in development and the information to be disclosed by Emergency Use Ventilator (EUV) manufacturers to the end user. These are based on the hazards identified in IEC 60601-1¹ and ISO 80601-2-80².

NOTE This document is intended to be used in conjunction with AAMI CR501:2020, *Emergency use ventilator (EUV) design guidance*.

1 Electrical Shock Hazard

Purpose: to ensure adequate patient and operator safety in terms of shock (leakage current, dielectric strength, ground continuity).

Disclosures:

- List AC input power requirements of the EUV (voltage, frequency, amperes).
 - DC power input requirement, if applicable.
 - Indicate the electrical classification of EUV:
 - Class I (EUV has a protective earth connection with a 3-wire power cord)
 - Class II (EUV does not have a protective earth ground but is double insulated with a 2-wire power cord)
 - Internally powered (powered by a rechargeable battery inside the EUV or external to EUV)
- Note An EUV can have more than one classification e.g., Class II/internally powered.
- If the power supply connected to mains power is not medical grade (i.e., IEC 60601-1 compliant), describe the means used to reduce leakage currents to IEC 60601-1 limits (e.g. use of an isolation transformer, second permanently installed protective earth connection).
 - If the power supply connected to mains power is Class I, add a warning:

¹ IEC 60601-1, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

² ISO 80601-2-80, *Medical electrical equipment - Part 2-80: Particular requirements for basic safety and essential performance of ventilatory support equipment for ventilatory insufficiency*

26 Warning: This ventilator relies on the integrity of the protective earth ground to reduce the risk of
27 electrical shock. Check the integrity and verify the function of the protective earth ground of the
28 supply mains receptacle prior to use.

- 29 • Describe the type of patient connection: basic, basic floating, cardiac floating (type B, BF or CF) and
30 defibrillation-proof.

31 **2 Mechanical Hazards**

- 32 a) Purpose: to ensure that the EUV can withstand mechanical stresses from being carried or wheeled
33 while being transported indoors or outdoors.

34 Disclosures:

- 35 • Identify the mobility of the EUV:
 - 36 ○ Transit operable: EUV is intended to operate while being moved.
 - 37 ○ Portable: EUV is intended to be carried (but not operating) from one location to another.
 - 38 ○ Mobile: EUV is intended to be wheeled (but not operating) from one location to another.
- 39 b) Purpose: to ensure that the moving parts of the EUV do not pose an unacceptable risk to the patient or
40 operator.

41 Disclosures:

- 42 • If the EUV has wheels, assess the stability and disclose the safe angle before tipping occurs.
- 43 • Identify any trapping zones (e.g. trapping fingers, hair, PPE) and how they are guarded.

44 **3 Environmental Hazards**

45 Purpose: to ensure that the EUV can be stored and operated in its intended environment.

46 Disclosures:

- 47 • Indicate the temperature/humidity/altitude range over which the EUV is intended to operate and
48 meets its specifications.
- 49 • Indicate the intended range of conditions (temperature/humidity specifications) in which the EUV
50 can be stored.

51 **4 Reuse Hazards**

52 Purpose: to reduce the risk of cross contamination.

53 Disclosures:

- 54 • Describe the cleaning and disinfection procedures needed between uses and between patients for
55 both the EUV and the accessories.
- 56 • Description of location and specifications of required EUV particle filters and replacement intervals.

57 **5 Biocompatibility**

58 Purpose: to reduce the risk of biological reaction to foreign substances.

59 Disclosures:

- 60 • For the gas pathway, indicate if any biocompatibility evaluations were performed per ISO 18562
61 (series)³.
- 62 • For parts intended to touch the patient, indicate if any biocompatibility evaluations were performed
63 per ISO 10993 (series)⁴.

64 **6 Electromagnetic Compatibility (EMC)**

65 Purpose: to ensure that the EUV is adequately protected from electromagnetic emissions from other
66 electrical sources (e.g. cell phones, ESD) and to ensure that the EUV does not interfere with the operation
67 of other nearby electronic medical devices.

68 Disclosures:

- 69 • Indicate if any EMC testing was performed and identify the standards (e.g., IEC 60601-1-2⁵) to
70 which the EUV was evaluated.
- 71 • If EMC testing has not been performed, add a warning:
72 This ventilator has not been tested for electromagnetic compatibility (EMC). It may produce
73 electromagnetic disturbances that will affect the performance of other equipment. It may fail to
74 perform as expected in the presence of electromagnetic disturbances from other equipment.

75 **7 Alarm System**

76 Purpose: to reduce the risk to the patient by alerting the caregiver of a hazardous situation.

77 Disclosures:

- 78 • Describe the functionality of the alarm system.
- 79 • List available alarm conditions, their relative priority and default alarm limits.
- 80 • Describe the visual alarm signals (e.g. text message) for each alarm condition.
- 81 • Describe the auditory alarm signals and how to discriminate between their priorities.
- 82 • Describe the default alarm settings (e.g. latched, not latched alarm signals, alarm condition
83 disabled).
- 84 • Indicate the means by which the auditory alarm signal can be inactivated and for how long.

³ ISO 18562, *Biocompatibility evaluation of breathing gas pathways in healthcare applications*

⁴ ISO 10993, *Biological evaluation of medical devices*

⁵ IEC 60601-1-2, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests*

85 **8 Accuracy of controls**

86 Purpose: to reduce the risk of hazardous output from the EUV to the patient.

87 Disclosures:

- 88 • List of displayed parameters: e.g., pressure, tidal volume, respiratory rate.
- 89 • Describe how the displayed parameters are measured or determined.
- 90 • List the accuracy of therapy parameters.

91 **9 Accessories**

92 Purpose: to ensure the safe use of the EUV with compatible accessories

93 Disclosures:

- 94 • List of recommended accessories and their replacement intervals e.g. tubing, patient interface,
95 filters, replacement batteries.

96 **10 Programmable Electrical Medical Systems**

97 Purpose: to ensure that the software operates safely and as specified.

98 Disclosures:

- 99 • Indicate whether the software was developed under a controlled life cycle process (e.g.,
100 IEC 62304⁶).
- 101 • List any known unresolved software anomalies and workarounds.
- 102 • Indicate:
103 Due to the rapid development cycle for this emergency use device, all efforts were made to verify
104 the software, but defects may still exist. The consequences of these defects are unknown and may
105 pose a risk to the patient.

106 **11 Risk Management Process**

107 Purpose: to ensure risks were comprehensively identified and adequately managed.

108 Disclosures:

- 109 • Indicate whether the EUV design has been developed using a risk management process (e.g.,
110 ISO 14971⁷).

111 **12 Other hazards**

112 Purpose: to reduce the risk of thermal injury or other events.

⁶ IEC 62304, *Medical device software — Software life cycle processes*

⁷ ISO 14971, *Medical devices - Application of risk management to medical devices*

113 Disclosures:

- 114 • If applicable, indicate the battery specifications including:
- 115 ○ the type of battery and chemistry;
- 116 ○ a description of the means to determine the status of the battery (e.g., charging, low battery
- 117 indicator);
- 118 ○ conformance to applicable standards (e.g., IEC 62133⁸ for rechargeable batteries or IEC
- 119 60086-4⁹ for non-rechargeable batteries).
- 120 • Indicate the ingress protection (IP) of the EUV enclosure: IP 22 is recommended (protection against
- 121 foreign objects \geq 12.5 mm and against dripping (15° tilted) water).
- 122 • Indicate if the EUV is suitable for use in an oxygen enriched environment $>$ 25 % O₂ (are adequate
- 123 protections in place to reduce risk of fire ignition).
- 124 • If the EUV contains oxygen at pressures exceeding 5 bar, the protections taken to ensure that auto-
- 125 ignition from adiabatic compression cannot occur (e.g., parts of the EUV operating at pipeline
- 126 pressure).

⁸ IEC 62133, *Secondary cells and batteries containing alkaline or other non-acid electrolytes – Safety requirements for portable sealed secondary cells, and for batteries made from them, for use in portable applications*

⁹ IEC 60086-4, *Primary batteries – Part 4: Safety of lithium batteries*